AD Enclosure - MT6 Section 2

Search and Secure Procedure

MT6 Section 2
MT6.2.V1.4



0) Note: Section 2 has a movable roof hatch, it may be open or closed during the secure.
00) If the Laser Power Supply is on, turn it off for the secure, then turn it back on when done.
1) "**A**" holds at the plastic curtain while:
   2) "**B**" resets Interlock Box #1, and checks the hallway.
   3) "**B**" searches the downstream section of the enclosure, using the mirror to check on top of the C-Block looking behind the Hut, then resets Interlock Box #2.
   4) "**B**" searches the Hut.
5) "**B**" joins "**A**" and together they search the upstream section of the enclosure and proceed to the entrance at the west labyrinth. If the Laser Power Supply was turned off for the secure, turn it on.
6) "**A**" and "**B**" exit via the west labyrinth, checking on top of the C-Blocks and the fence by the Portakamp.
7) "**A**" and "**B**" exit and reset Interlock Box #3.

Reviewed AD/RSO	Date
Gary Lauten	3/28/14
Approved AD/OPS Dept. Head	Date
Daniel Johnson	3/28/14

CONTROLLED DOCUMENT
Users are responsible for ensuring they work to the latest approved revision. Printed or electronically transmitted copies are uncontrolled.